

Declaration of Conformity



Sartorius Biohit Liquid Handling Oy hereby certifies that:

Andrew Alliance Picus Pipette

Conforms to the essential requirements of **Directive 2014/30/53 EU** and protection requirements of:

- Electromagnetic Compatibility, **EN 61326-1:2013** with EMC Test Report no. 03012018_005, date : 7th February 2019 issued by: Espotel Oy, Mattilanniemi 6-8, FI-40100 Jyväskylä, p. +358 9 8565 8600
- Electrical Safety **IEC 61010-1:2010/AM1 (Ed.3.1) and EN 61010-1:2010/AM1:2016(Draft)** with Test Report no. 1-7653/18-01-03-A date: 21st March 2019, issued by: CTC advanced GmbH, Unterturkheimerstr. 6-10, D-66117, Saarbrücken, Germany
- Radio Frequency Compatibility, **EN 300 328, v.2.1.1** with RF Test Report no. 03012018_006, date : 7th February 2019 issued by: Espotel Oy, Mattilanniemi 6-8, FI-40100 Jyväskylä, p. +358 9 8565 8600
- Radio Frequency and emission Compatibility, **FCC-Title 47 CFR, Part 15:2019-03 and ANSI C63.4:2014** with RF Test Report no. 1-7653/18-01-04, issued by: CTC advanced GmbH, Unterturkheimerstr. 6-10, D-66117, Saarbrücken, Germany
- RoHS Directive **2011/65/EU**.

This declaration is supported by following EC quality system approval certificates:

ISO 9001:2015 Certificate No. FI17/5101, issued by SGS, Finland, valid 14th December 2020

ISO 13485:2016 Certificate No. FI17/5103, issued by SGS, Finland, valid 14th December 2020

ISO 17025:2005 Certificate No. KO41/A10/2016 issued by FINAS, valid 20th March 2024

ISO 14001:2015 Certificate No. FI17/5102, issued by SGS, Finland, valid 14th December 2020

26th June 2020

Seppo Riikonen

Director, Quality & Process Development

Sartorius Biohit Liquid Handling Oy